



校准结果

RESULTS OF CALIBRATION

证书编号 H190330005001
Certificate No.

原始记录号 190330005001
Record No.

第 6 页 共 7 页
Page of

中心点实测平均值 Center point measured average	显示值的平均值 Average value of displayed values	误差 Error	允许误差 Allowable error	结论 in conclusion
84.89 °C	85.01 °C	0.12 °C	±2 °C	PASS
84.0 %RH	85.0 %RH	1.0 %RH	±5 %RH	PASS
温度均匀度 Temperature uniformity	温度波动度 Temperature fluctuation	湿度均匀度 Humidity uniformity	湿度波动度 Humidity fluctuation	
0.54 °C	± 0.02 °C	0.2 %RH	± 0.1 %RH	

温度校准点 150 °C

Temperature calibration point 150 °C

下表中，温度测试位置为1~9，实测值单位为°C

In the table below, the temperature test position is 1~9, and the measured value is °C.

测试位置 Test location	1	2	3	4	5	6	7	8	9
实测最大值 Measured maximum	151.18	150.63	151.29	151.22	151.28	150.92	150.68	150.42	150.52
实测最小值 Measured minimum	151.13	150.58	151.24	151.18	151.24	150.87	150.63	150.38	150.48

中心点实测平均值 Center point measured average	显示值的平均值 Average value of displayed values	误差 Error	允许误差 Allowable error	结论 in conclusion
150.50 °C	150.01 °C	-0.49 °C	±2 °C	PASS
温度均匀度 Temperature uniformity	温度波动度 Temperature fluctuation			
0.86 °C	± 0.02 °C			

