


## 4. Run operation settings



[Figure 4-1] Run operation set screen

- ① Select a setting from the Program and Fix.
- ② Power up operation when setting power cut.
  - ☞ STOP : After the operation is interrupted, the state is restored to the operation of the stop state when the power is restored.
  - ☞ COLD: After the operation is interrupted, the operation is restarted when the power is restored.
  - ☞ HOT : After the operation is interrupted, the state can be restored to the state before the power is off when the power is restored.
- ③ When locked, all parameters are set to be not used.
- ④ When changing the set value, it can increase or decrease automatically according to the proportion in advance. (only applicable to set values at run time).
- ⑤ The set time and the [Figure 3.2.3 Fix Run second run screen] in the set of the overall running time setting and running the termination of the operation. (only applicable to Fix at run time).

Enter the lock set enter password: default is 0

**LOCK SET**

2019-11-13  
15 : 56 : 27

 **LOCK SET**

PASSWORD

 **AUTO LOCK**

OFF  ON

 **BACK**

